

## EU DECLARATION OF CONFORMITY

Model: **TOPAS Sonic M**  
Type designation: **TSM**

We,

**INTEGRA METERING**  
12 rue Font Grasse  
31700 Blagnac, France

declare under our sole responsibility the product listed above consistent with the type described in the certificate No. **SK22-MI001-SMU072** – revision 2 (issued February 26<sup>th</sup>, 2025 (valid until February 17<sup>th</sup>, 2032) and approved by the Notified body 1781 Slovak Institute of Metrology in accordance with **Directive 2014/32/EU Annex II, Module B**).

The quality system for production, final product inspection and testing of the **water meters (MI-001)** was approved by the Notified body 1781 Slovak Institute of Metrology in accordance with **Directive 2014/32/EU Annex II, Module D**, under Certificate No. **SK 22-QD-SMU020-revision 5** (issued July 2<sup>nd</sup>, 2024 and valid until August 19<sup>th</sup>, 2025).


The object of the declaration described above is in conformity with the essential requirements of:


- **Directive 2014/32/EU** on Measuring instrument
- **Directive 2014/30/EU** on Electromagnetic compatibility
- **Directive 2014/53/EU** on Radio equipment
- **Directive 2014/35/EU** on Low voltage electrical equipment.
- **Directive (EU) 2015/863 of 31 March 2015 amending Annex II to Directive 2011/65/EU** on restriction of use of certain hazardous substances in electrical and electronic equipment (RoHS III)
- **Directive 1907/2006** on Registration, Evaluation, Authorization and Restriction of Chemical (REACH)

and harmonized standards and normative documents

- **OIML-R49-1/2/3** *Water meters intended for the metering of cold potable water - Part 1: Metrological and technical requirements (2006) / Part 2: Test methods (2004-2013) / Part 3: Test Report Format (2013)*
- **ISO 4064-1/2/3: 2014** *Water meters for cold potable water and hot water - Part 1: Metrological and technical requirements / Part 2: Test methods / Part 3: Test report format*
- **IEC 61010-1:2010+A1: Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements**
- **ETSI EN 300 220-1/2:2017** : *Short Range Devices (SRD) operating in the frequency range 25 MHz to 1 000 MHz*
- **ETSI EN 300 330:2017** : *Short Range Devices (SRD); Radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz*
- **EN 61326-1:2013** : *Electrical equipment for measurement, control and laboratory use - EMC requirements*
- **EN 62311:2008** : *Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz - 300 GHz)*
- **EN 50364:2010** : *Limitation of human exposure to electromagnetic fields from devices operating in the frequency range 0 Hz to 300 GHz, used in Electronic Article Surveillance (EAS), Radio Frequency Identification (RFID) and similar applications*
- **ETSI EN 301 489-1/3:2019** : *Electromagnetic Compatibility (EMC) standard for radio equipment and services*

Blagnac, 27/02/2025

  
Mathieu POQUE  
Head of Quality

  
John ARNOLD  
CEO